LH Ovulation Rapid Test

Qualitative LH Surge Detection

NIP-RT-FRT-002

- Rapid diagnostic test datasheet



PRODUCT OVERVIEW

Rapid qualitative urine test for detection of the luteinizing hormone surge to help identify the period when ovulation is likely to occur.

SPECIMEN	FORMAT	RESULT	STORAGE
Urine	Cassette	5 min	4-30 °C

SHELF LIFE	ANALYTICAL SENSITIVITY
36 months	25 mIU/mL

COMMERCIAL NOTE

Final packaging configuration, instructions for use and market-specific claims must follow the approved Nipigon product labeling and applicable authorization.

Technical Information

LH Ovulation Rapid Test

INTENDED USE

Qualitative detection of an LH surge in urine as an aid in predicting ovulation.

USE CLASSIFICATION

For in vitro self-testing where permitted by the approved labeling and local market authorization.

Nipigon Catalog No.	**NIP-RT-FRT-002**
Specimen	**Urine**
Format	**Cassette**
Time to Result	**5 min**
Storage	**4-30 °C**
Shelf Life	**36 months**
Detection Target	**Luteinizing hormone (LH)**

IMPORTANT

This datasheet summarizes the selected Nipigon configuration. Final instructions, interpretation, warnings, pack size and regulatory statements must follow the approved product IFU and labeling for the destination market.