
Sperm Concentration Rapid Test

Qualitative Semen Concentration Screening

NIP-RT-FRT-004

Rapid diagnostic test datasheet



PRODUCT OVERVIEW

Rapid qualitative test for assessment of sperm concentration in human semen as an aid in evaluating male fertility status.

SPECIMEN	FORMAT	RESULT	STORAGE
Semen	Cassette	5-10 min	4-30 °C

SHELF LIFE	DETECTION LIMIT
24 months	15 × 10^6/mL

COMMERCIAL NOTE

Final packaging configuration, instructions for use and market-specific claims must follow the approved Nipigon product labeling and applicable authorization.

Technical Information

Sperm Concentration Rapid Test

INTENDED USE

Qualitative detection of sperm concentration in human semen relative to the stated detection threshold.

USE CLASSIFICATION

For in vitro diagnostic self-testing where permitted by the approved labeling and local market authorization.

Nipigon Catalog No.	**NIP-RT-FRT-004**
Specimen	**Semen**
Format	**Cassette**
Time to Result	**5-10 min**
Storage	**4-30 °C**
Shelf Life	**24 months**
Detection Target	**Sperm concentration threshold**

IMPORTANT

This datasheet summarizes the selected Nipigon configuration. Final instructions, interpretation, warnings, pack size and regulatory statements must follow the approved product IFU and labeling for the destination market.