Chlamydia Rapid Test

Chlamydia trachomatis Antigen Detection

NIP-RT-INF-012

Professional rapid diagnostic test datasheet



PRODUCT OVERVIEW

Rapid immunoassay for direct qualitative detection of Chlamydia trachomatis antigen in female endocervical swab, male penile swab and male urine specimens for professional in vitro diagnostic use.

SPECIMEN	FORMAT	RESULT	STORAGE
Swab / Male Urine	Cassette	10-15 min	4-30 °C

SHELF LIFE	DETECTION TARGET
See approved labeling	Chlamydia trachomatis antigen

Technical Information

Chlamydia Rapid Test

INTENDED USE

Rapid immunoassay for direct qualitative detection of Chlamydia trachomatis antigen in female endocervical swab, male penile swab and male urine specimens for professional in vitro diagnostic use.

PROFESSIONAL USE

For professional in vitro diagnostic screening. Reactive screening results require follow-up according to the applicable testing algorithm and approved IFU.

Nipigon Catalog No.	NIP-RT-INF-012
Specimen	Swab / Male Urine
Format	Cassette
Time to Result	10-15 min
Storage	4-30 °C
Shelf Life	See approved labeling
Detection Target	Chlamydia trachomatis antigen
Sample Volume	Specimen extraction per approved IFU

PRODUCT NOTE

For professional in vitro diagnostic use. Use the approved market-specific labeling and instructions for use for specimen handling, test procedure, result interpretation and regulatory information.